

GasAlert EXTREME

Single Gas Detector

User Manual



Limited Warranty and Limitation Liability

BW Technologies LP (BW) warrants the product to be free from defects in material and workmanship under normal use and service for a period of two years, beginning on the date of shipment to the buyer. This warranty extends only to the sale of new and unused products to the original buyer. BW's warranty obligation is limited, at BW's option, to refund of the purchase price, repair or replacement of a defective product that is returned to a BW authorized service center within the warranty period. In no event shall BW's liability hereunder exceed the purchase price actually paid by the buyer for the Product.

This warranty does not include:

- a) fuses, disposable batteries or the routine replacement of parts due to the normal wear and tear of the product arising from use;
- b) any product which in BW's opinion, has been misused, altered, neglected or damaged, by accident or abnormal conditions of operation, handling or use;
- c) any damage or defects attributable to repair of the product by any person other than an authorized dealer, or the installation of unapproved parts on the product; or The obligations set forth in this warranty are conditional on:
 - a) proper storage, installation, calibration, use, maintenance and compliance with the product manual instructions and any other applicable recommendations of BW;
 - b) the buyer promptly notifying BW of any defect and, if required, promptly making the product available for correction. No goods shall be returned to BW until receipt by the buyer of shipping instructions from BW; and
- c) the right of BW to require that the buyer provide proof of purchase such as the original invoice, bill of sale or packing slip to establish that the product is within the warranty period. THE BUYER AGREES THAT THIS WARRANTY IS THE BUYER'S SOLE AND EXCLUSIVE REMEDY AND IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. BW SHALL NOT BE LIABLE FOR ANY SPECIAL, INDIRECT, INCIDENTAL, OR BASED ON CONTRACT, TORT OR RELIANCE OR ANY OTHER THEORY.

Since some countries or states do not allow limitation of the term of an implied warranty, or exclusion or limitation of incidental or consequential damages, the limitations and exclusions of this warranty may not apply to every buyer. If any provision of this warranty is held invalid or unenforceable by a court of competent jurisdiction, such holding will not affect the validity or enforceability of any other provision.

Table of Contents

Title	Page
Introduction	1
Contacting BW Technologies by Honeywell	
Safety Information - Read First	
Getting Started	
Activating the Detector	9
Self-Test	9
Self-Test Pass	12
Self-Test Fail	12
Deactivating the Detector	
Confidence Beep	
User Options Menu	
Exit	
Clock Option	
Passcode Protection Option	
Enable Passcode Protection	
Disable Passcode Protection	
Deactivation Passcode Protection	
Stealth Mode Option	18
Automatic Backlight Option	
Latching Alarm Option	
Automatic Oxygen (O ₂) Calibration Option	20
Calibration Past Due Option	21
Languages	21

User Manual

Title	Page
Portuguese Option	21
Spanish Option	
German Option	22
French Option	22
English Option	23
Datalogger Sampling Rate Option	23
Data Transfer Option	24
LAST and ALL Transfers	25
EVNT Transfer	25
Unsuccessful Transfer	25
Alarms	26
Computed Gas Exposures	28
Viewing Gas Exposures	28
Toxic Gases	28
Oxygen	28
Gas Alarm Setpoints	29
Stopping a Gas Alarm	29
Clearing Gas Exposures	
Resetting Gas Alarm Setpoints	
Sensor Alarm	
Low Battery Alarm	
Automatic Shutdown Alarm	
Calibration and Setting Alarm Setpoints	
Guidelines	
Test Cap	
Calibration	
Start Calibration	34

Title	Page
Auto Zero	34
Auto Zero Fail	34
Passcode Protected	35
Set Span	35
Span	
Setting the Calibration Due Date	37
Setting the TWA Alarm Setpoint	
Setting the STEL Alarm Setpoint	
Setting the Low Alarm Setpoint	
Setting the High Alarm Setpoint	
Verification	
Datalog and Event Log	
Datalog	
Event Log	
Maintenance	
Replacing the Battery or Sensor	
Replacing the Battery	
Replacing the Sensor	
Cleaning a Sensor Screen	
Clearing a Sensor	
WEEE Directive and Battery Directive	
Removal and Disposal of the Battery	
Troubleshooting	
Replacement Parts and Accessories	
Specifications	
General Specifications for Datalogger Units	

User Manual

List of Figures

Title	Figure	Page
1.	GasAlert Extreme Detector	6
2.	Display Elements	7
3.	Test Cap	
4.	Replacing the Battery or Sensor	41
5.	Rear Shell Seal	42
6.	Front Shell Seal	42

User Manual

List of Tables

Table	Title	Page
1.	GasAlert Extreme Models	1
2.	International Symbols	
3.	GasAlert Extreme Detector	
4.	Display Elements	7
5.	Pushbuttons	8
6.	Alarms	26
7.	Computed Gas Exposures	
8.	Gas Alarm Setpoints	29
9.	Factory Alarm Setpoints	30
10.	Test Cap	33
11.	Replacing the Battery or Sensor	41
12.	Rear Shell Seal	
13.	Front Shell Seal	42
14.	Troubleshooting Tips	
15	Replacement Parts and Accessories	

User Manual

Introduction

Marning

To ensure your personal safety, read "Safety Information" before using the detector.

The GasAlert Extreme gas detector ("the detector") warns of hazardous gas at levels above a factory set alarm setpoint. This product is a gas detector, not a measurement device.

The detector is a personal safety device. It is your responsibility to respond properly to the alarms.

Table 1 lists the GasAlert Extreme models and the gases monitored. This manual includes examples from each model.

Table 1. GasAlert Extreme Models

Model	Gas Monitored
GasAlert Extreme O ₂	Oxygen (% by volume)
GasAlert Extreme CO	Carbon monoxide (ppm) Low H ₂ sensitivity
GasAlert Extreme CO	Carbon monoxide (ppm)
GasAlert Extreme H ₂ S	Hydrogen sulfide (ppm) High range

Model	Gas Monitored
GasAlert Extreme H ₂ S	Hydrogen sulfide (ppm)
GasAlert Extreme PH ₃	Phosphine (ppm)
GasAlert Extreme SO ₂	Sulfur dioxide (ppm)
GasAlert Extreme Cl ₂	Chlorine (ppm)
GasAlert Extreme NH ₃	Ammonia (ppm)
GasAlert Extreme NH ₃	Ammonia (ppm) High range
GasAlert Extreme NO ₂	Nitrogen dioxide (ppm)
GasAlert Extreme HCN	Hydrogen cyanide (ppm)
GasAlert Extreme ETO	Ethylene oxide (ppm)
GasAlert Extreme ClO ₂	Chlorine dioxide (ppm)
GasAlert Extreme O ₃	Ozone (ppm)
GasAlert Extreme NO	Nitric oxide (ppm)

Safety Information - Read First

Use the detector only as specified in this manual, otherwise the protection provided by the detector may be impaired.

International symbols used on the detector and in this manual are explained in Table 2.

Read the **Warnings** and **Cautions** on the following pages before using the detector.



This instrument contains a lithium battery. Do not mix with the solid waste stream. Spent batteries should be disposed of by a qualified recycler or hazardous materials handler. User Manual

⚠ Caution

To avoid possible personal injury, adhere to the following:

- · Warning: Substitution of components may impair Intrinsic Safety.
- Warning: To prevent ignition of flammable or combustible atmospheres, disconnect power before servicing.
- Do not use the detector if it is damaged. Before using the detector, inspect the case. Look for cracks and missing parts.
- Use only a sensor specifically designed for the GasAlert Extreme model. Refer to Replacement Parts and Accessories.
- Do not deactivate the detector during a work shift. Deactivating the detector resets the time-weighted average (TWA), short-term exposure limit (STEL), and maximum gas exposure values to 0. Refer to <u>Alarms</u>.
- Ensure the sensor screen is not blocked.
- If the detector is damaged or parts are missing, contact <u>BW Technologies by Honeywell</u> immediately.
- If the detector has been disassembled, ensure the front and rear shells are properly aligned and fastened before activating the detector. Refer to Maintenance.
- BW recommends to bump test the sensor, before each day's use, to confirm their ability to respond to gas by exposing the detector to a gas concentration that exceeds the alarm setpoints. Manually verify that the audible and visual alarms are activated. For detectors with the calibration feature, calibrate the sensor if the reading is not within the specified limits.
- Calibrate the detector before first-time use, and then at least once every 180 days. For HCN detectors, calibrate once
 every 30 days.
- Use only the Panasonic CR-2 battery. Refer to Replacing the Battery or Sensor.
- To reduce the risk of ignition of a flammable atmosphere, batteries must only be changed in a safe area free of hazardous gas.

▲ Caution

To avoid possible damage to the detector, adhere to the following:

- Do not expose the detector to electrical shock and/or severe continuous mechanical shock.
- The oxygen GasAlert Extreme detector is classified by Underwriters Laboratories Inc. up to an atmosphere of 21% oxygen.
- Do not attempt to disassemble, adjust, or service the detector unless instructions for that procedure are contained in the user manual, and/or that part is listed as a replacement part. Use only BW Technologies by Honeywell replacement parts. Refer to Replacement Parts and Accessories.
- The detector warranty will be voided if customer personnel or third parties damage the detector during repair attempts. Non-BW Technologies by Honeywell repair/service attempts void this warranty.

User Manual

∧ Attention

Pour éviter toute blessure possible, respectez ce qui suit :

- Avertissement : Le remplacement d'un composant de l'appareil peut altérer sa sécurité intrinsèque.
- Avertissement : Pour éviter l'inflammation d'atmosphères inflammables ou combustibles, coupez l'alimentation électrique avant tout entretien.
- N'utilisez pas le détecteur s'il est endommagé. Avant d'utiliser le détecteur, inspectez le boîtier. Assurez-vous qu'aucune pièce ne manque et vérifiez l'absence de fissures.
- Utilisez uniquement un capteur spécialement conçu pour le modèle GasAlert Extreme. Reportez-vous à la section Replacement Parts and Accessories.
- Ne désactivez pas le détecteur pendant un quart de travail. Si vous désactivez le détecteur, les valeurs de la moyenne pondérée dans le temps (MPT), de la limite d'exposition (LECT) à court terme et d'exposition maximale aux gaz sont réinitialisées à 0. Reportez-vous à la section Alarms.
- · Assurez-vous que l'écran du capteur est verrouillé.
- Si le détecteur est endommagé ou que des pièces sont manquantes, contactez immédiatement BW Technologies by Honeywell.
- Si le détecteur a été démonté, assurez-vous que les coques avant et arrière sont correctement alignées et serrées avant d'activer le détecteur. Reportez-vous à la section Maintenance.
- Avant chaque utilisation quotidienne, BW recommande d'effectuer un test fonctionnel du capteur afin de vérifier qu'il réagit bien aux gaz présents, en exposant le détecteur à une concentration de gaz supérieure aux seuils d'alarme. Vérifiez manuellement que les alarmes sonore et visuelle sont activées. Pour les détecteurs dotés de la fonction d'étalonnage, étalonnez le capteur si la valeur ne se trouve pas dans les limites spécifiées.
- Étalonnez le détecteur lors de la première utilisation, puis une fois tous les 180 jours. Pour les détecteurs HCN, étalonnez-les tous les 30 jours.
- Utilisez uniquement la batterie Panasonic CR-2. Reportez-vous à la section Replacing the Battery or Sensor.
- Afin de réduire le risque d'inflammation d'une atmosphère inflammable, les batteries doivent être changées dans un lieu sûr exempt de gaz dangereux.

⚠ Attention

Pour éviter d'endommager le détecteur, respectez ce qui suit :

- Ne soumettez pas le détecteur à des chocs électriques et/ou à d'importants chocs mécaniques continus.
- Le détecteur d'oxygène GasAlert Extreme est classifié conforme par Underwriters Laboratories Inc. jusqu'à une atmosphère contenant 21 % d'oxygène.
- N'essayez pas de démonter, d'ajuster ou d'entretenir le détecteur, sauf si les instructions pour cette procédure vous sont fournies dans le manuel d'utilisation et/ou si la pièce concernée est répertoriée comme pièce de rechange. Utilisez uniquement des pièces de rechange de BW Technologies by Honeywell. Reportez-vous à la section Replacement Parts and Accessories.
- La garantie du détecteur sera annulée si le personnel du client ou des tiers ont endommagé le détecteur pendant des tentatives de réparation. Les tentatives de réparation/d'entretien effectuées par des parties autres que BW Technologies by Honeywell entraîneront l'annulation de cette garantie.

Table 2. International Symbols

Symbol	Meaning	
C UL US	Classified to both U.S. and Canadian Safety standards by Underwriter's Laboratories, Inc.	
CE	Conforms to European Union Directives	
⟨£x⟩	European Explosives Protection	
ATEX	Conforms to European ATEX Directives	
IECEx	International Electrotechnical Commission Scheme for Certification to Standards for Electrical Equipment for Explosive Atmospheres	
EAC Ex	Conforms to Russian Custom Union Certification and Declaration	
K s	S Conforms to Korea Testing Laboratory (KTL) Certification	
INMETRO	Conforms to Brazilian InMetro Certification	
	Australian Regulatory Compliance Mark (RCM)	

Getting Started

The items listed below are included with the detector. If the detector is damaged or parts are missing, contact the place of purchase immediately.

- 3 V lithium CR2-series battery
- GasAlert Extreme O₂ model: O₂ sensor;
 GasAlert Extreme CO model: CO sensor

(low H₂ sensitivity);

GasAlert Extreme CO model: CO sensor;

GasAlert Extreme H₂S model: H₂S sensor (high range);

GasAlert Extreme H₂S model: H₂S sensor; GasAlert Extreme PH₃ model: PH₃ sensor; GasAlert Extreme SO₂ model: SO₂ sensor; GasAlert Extreme CI₂ model: CI₂ sensor; GasAlert Extreme NH₃ model: NH₃ sensor:

GasAlert Extreme NH₃ model: NH₃ sensor (high range);

GasAlert Extreme NO₂ model: NO₂ sensor; GasAlert Extreme HCN model: HCN sensor;

GasAlert Extreme ETO model: ETO sensor; GasAlert Extreme CIO₂ model: CIO₂ sensor;

GasAlert Extreme O₃ model: O₃ sensor;

GasAlert Extreme NO model: NO sensor.

Test cap and hose

The detector is shipped with the battery and sensor installed. To order replacement sensors and accessories, refer to Replacement Parts and Accessories.

To become familiar with the features and functions of the detector, study the following figures and tables:

- Figure 1 and Table 3: GasAlert Extreme Detector (describes the detector's components).
- Figure 2 and Table 4: Display Elements (describes the LCD screen and icons).
- Table 5: Pushbuttons (describes the buttons on the detector.




Figure 1. GasAlert Extreme Detector

Table 3. GasAlert Extreme Detector

Item	Description
1	Visual alarm indicators (LEDs)
2	Liquid crystal display (LCD)
3	Pushbuttons
4	Audible alarm
5	Sensor and sensor screen
6	Infrared (IR) communication port
7	Alligator clip




Figure 2. Display Elements

Note

When enabled, the backlight option automatically activates for 3 seconds whenever there is insufficient light to view the LCD. Press and hold (until the backlight activates) any button to activate the backlight for 6 seconds. The detector is shipped with the backlight option enabled.

The backlight does not operate when stealth mode is enabled.

Table 4. Display Elements

Item	Description
1	Numeric value
2	Gas cylinder
3	Automatically span sensor
4	Passcode lock
5	Set alarm setpoints and user options
6	Maximum gas exposure
7	Alarm conditions
8	Battery
9	Data transmission
10	Alarm or alarm setpoint
11	Automatically zero sensor
12	Optional datalogger indicator
13	Parts per million (ppm)
14	Percentage by volume (% vol.)
15	Percentage by lower explosive limit (% LEL) (future use)

Table 5. Pushbuttons

Pushbutton	Description	
©	To activate the detector, press	
	• To enable/disable the confidence beep, while the detector is deactivated press and hold \(\). While holding \(\), press \(\emptyre{0} \) to enable or disable the confidence beep during start-up.	
	 To deactivate the detector, press	
▼	To decrement the displayed value or to scroll down, press ▼.	
	 To enter the user options menu, press ▼ and ▲ simultaneously and hold until OPTN and then EXIT displays (5 seconds). 	
	 To initiate calibration and set alarm setpoints, press and hold	
A	To increment the displayed value, press .	
	To view the TWA, STEL and maximum (MAX) gas exposures, press and simultaneously.	
	To save a displayed value, press	
\bigcirc	To clear TWA, STEL, and maximum (MAX) gas exposures, press and hold for 6 seconds.	
	To acknowledge a latched alarm, press	

Activating the Detector

To activate the detector, press ① in a safe area, free of hazardous gas.

Self-Test

When the detector is activated, it performs several self-tests. Confirm the following tests occur.

Note

The following tests are listed in the order they are automatically performed on the detector.

1. **Display Elements Test:** The LCD displays all screen elements.



- 2. **Alarm Function Test:** The detector beeps, the LEDs flash, the backlight activates briefly, and the detector vibrates.
- Battery Test: The detector tests the batteries. If the battery voltage is too low to continue, the detector performs an automatic shutdown. Refer to <u>Automatic Shutdown Alarm</u>.

 Date and Time: The LCD displays the date and time automatically in the following order.

Set	Year: The LCD displays the current year (20XX).
Set	Month: JAN, FEB, MAR, etc.
1	Day of the month: (1 to 31)
MEIN	Day of the week: MON, TUE, WED, etc.
	Hour/Minute: 00:00 hours to 23:59 hours

To adjust the date or time, refer to Clock Option.

 Sensor Test: The detector tests the sensor. If the sensor test fails, the detector beeps slowly, the LEDs flash slowly, and ALARM flashes.



If the sensor test passes, the self-test continues.

Gas Type: The LCD displays the type of gas the detector is manufactured to monitor.



Refer to Table 1 for the type of gases monitored.

If the battery is low, the LCD displays the low battery icon and the self-test continues.



TWA Alarm Setpoint: The LCD displays the TWA alarm setpoint.



Note

The TWA alarm setpoint screen does not apply to O_2 detectors.

STEL Alarm Setpoint: The LCD displays the STEL alarm setpoint.



Note

The STEL alarm setpoint screen does not apply to O_2 detectors.

9. Low Alarm Setpoint: The LCD displays the low alarm setpoint.



10. High Alarm Setpoint: The LCD displays the high alarm setpoint.




 Calibration Due Test: The LCD displays the calibration due date



The LCD displays the number of days remaining before the detector must be calibrated. For more information, refer to <u>Calibration</u>.

If calibration is past due, CAL. PAST displays.



Press to acknowledge the warning message.

If the calibration past due option is enabled or disabled, one of the following two events will occur.

- Calibration Past Due Disabled: If the detector is not passcode protected and after the CAL. PAST message is acknowledged, the detector continues the self-test and then enters normal operation.
- Calibration Past Due Enabled: If the detector is passcode
 protected and when CAL. PAST displays, press
 to
 acknowledge the message and to access the PASS screen. If
 required, refer to Passcode Protection Option.



Press • or • to scroll to the required passcode, and press • within 10 seconds to confirm the selection. The detector enters normal operation.

Note


Calibrate the detector before continuing operation.

If the passcode is not confirmed within 10 seconds or the passcode is incorrect, the detector beeps eight times, the LEDs flash 8 times, and the LCD displays the following screen.



The detector then automatically deactivates.

12. **Bump Test Fail:** If a previous bump test failed, the detector beeps, vibrates, and **BUMP FAIL** displays:.



Press () to acknowledge the alarm.

Note

Bump test the detector before continuing operation.

For information regarding bump tests, refer to the *MicroDock II User Manual*.

Self-Test Pass

If the detector passes the self-test, it enters normal operation. The LCD displays the ambient gas reading.



The detector begins recording immediately. It records the

- maximum (MAX) gas exposure,
- the short-term exposure levels (STEL), and
- · calculates the time-weighted average (TWA).

Self-Test Fail

If the detector fails the self-test, refer to <u>Troubleshooting</u>.

Deactivating the Detector

Note

If Deactivation Passcode Protection is enabled, the detector cannot be disabled without entering a passcode first. PASS displays immediately after OFF, refer to <u>Deactivation Passcode Protection</u>. This option can only be enabled at the factory and cannot be disabled by the customer.

To deactivate the detector, complete the following:

1. Press and hold ① until **OFF** displays (approximately 5 seconds).



The detector beeps and vibrates four times, the LEDs flash four times, and then the detector deactivates.

Note

If (a) is not held down until OFF displays, the detector will remain activated.

Confidence Beep

The confidence beep provides continuous confirmation that the detector is operating properly. When the confidence beep is enabled, the detector beeps every 5 seconds.

The confidence beep can be enabled or disabled during start-up.

Note

The detector is shipped with the confidence beep disabled.

Enabling the confidence beep decreases battery life.

To enable/disable the confidence beep, complete the following:

- 1. Ensure the detector is deactivated.
- 2. Press and hold . While holding , press .

When the confidence beep option is enabled, the detector automatically begins beeping when activated.

When the confidence beep option is enabled in stealth mode, the detector vibrates one time every 60 seconds. For more information refer to <u>Stealth Mode Option</u> and <u>Alarms</u>.

Note

If confidence beep is enabled and a low battery alarm occurs, the confidence beep deactivates.

GasAlert Extreme User Manual

User Options Menu

Note

When selecting a user option, **Set** flashes and the LCD displays the opposite of what is currently enabled.



To access the user options menu, complete the following:

 Press and hold and simultaneously until OPTN displays and then release the buttons.

The detector beeps and vibrates four times and the LEDs flash four times while accessing the user options menu.

If the passcode protection is not enabled, the **EXIT** screen automatically displays.



If the detector is passcode protected, the following screen displays.



2. Press ▲ or ▼ to scroll to the required passcode. Press ○ to confirm the selection and access the **EXIT** screen.

Note

If the passcode is not confirmed within 10 seconds, **NO** displays and the detector returns to normal operation.



- 3. From the **EXIT** screen, press ▲ or ▼ to scroll through the user options.
- 4. Press () to select a displayed option.

Note

As a safety precaution, if an option is not selected within 20 seconds the detector automatically returns to normal operation.

When the required activities have been performed for a selected option, the EXIT screen automatically displays.

5. Press ♠ or ▼ to select another option or press ○ to exit the user options menu and return to normal operation.

User Manual

Exit

When entering user options, the **EXIT** screen displays immediately following the options (**OPTN**) screen. The LCD automatically returns to the **EXIT** screen after a user option has been accessed.

From the **EXIT** screen, use rianlge or rianlge to scroll to additional user options. Or

Press O to exit user options and return to normal operation.

Clock Option

The clock (**CLCK**) option sets the date (year/month/day/day of the week) and time (hour/minute) of the detector. To set the time or date, complete the following:

- From the EXIT screen, press ▲ or ▼ to scroll to the CLCK option.
- 2. Press to select the option and access the first date/time option, the year. **Set** and the last two digits of the year continually flash.
- 3. Press ▲ or ▼ to scroll to the required year and press to confirm the selection.

Or

To bypass the year, press O to retain the current value and automatically proceed to the month screen.

4. Repeat step #3 for the remaining date/time changes.

5. Press ▲ or ▼ to select another option or press ○ to exit the user options menu and return to normal operation.

Note

The time and date values can only be changed in the order they are presented in this table. To bypass any setting, press . The detector automatically retains the current value and proceeds to the next date/time option.

Set	Year: Requires only the last two numerals of the year (00-99).
Set	Month: Scroll to select the required month (JAN, FEB, MAR, etc.).
Set C. J.	Day: Scroll to select the required day (1-31). For months with 30 days (1-30) is available. For February, (1-28 & 29) is available.
Set NAITINI	Day of the week: Scroll to select the required day (MON, TUE, WED, etc.).
Set	Time: The hour value flashes first. Scroll to select (0:00 hours. to 23:59 hours).

User Manual

Note

If a value is not bypassed by pressing within 10 seconds, the detector automatically proceeds to the next date/time option. If the Time Minute value was not bypassed, the detector automatically proceeds to the Exit screen.

If a new value is selected but not confirmed by pressing within 10 seconds, NO displays and the detector proceeds to the next date/time option. If a new Time Minute value was selected but not confirmed, the detector automatically proceeds to the Exit screen.



Passcode Protection Option

The passcode protection option (**PASS**) prevents unauthorized access to the user options and the calibration/set alarm setpoint functions.

The passcode protection option can be enabled or disabled.

Note

The detector is shipped with the passcode protection option disabled.

Enable Passcode Protection

To enable passcode protection, complete the following:

Note

The passcode is provided on a separate card inside the shipping container.

From the EXIT screen of the user options menu, press ▲ or ▼
to scroll to the PASS option.



- 2. Press () to select the option.
- 3. **Set** and **PASS** continually flash. Press ▲ or ▼ to scroll to the required passcode, and press to confirm the selection.
- 4. The **ON** screen displays and flashes continually. Press to confirm. The LCD then returns to the **EXIT** screen.



5. Press ▲ or ▼ to select another user option, or press ○ to exit the user options and return to normal operation.

Note

If an incorrect passcode is selected or a correct passcode is not confirmed within 10 seconds, **NO** displays and the LCD returns to the **EXIT** screen.



User Manual

Disable Passcode Protection

When the detector is passcode protected, the key icon displays continually. To disable the passcode protection option, complete the following:

 Press and hold ▲ and ▼ simultaneously to access the user options menu.

The **OPTN** screen displays briefly before the flashing passcode screen displays.



2. Press ▲ or ▼ to scroll to the required passcode and press ○ to confirm. The following **EXIT** screen displays.



The key icon indicates that the passcode protection is currently enabled.

- 3. Press ▲ or ▼ to scroll to the **PASS** option, and press to select the option.
- 4. The LCD displays a flashing **OFF** screen. Press to confirm the disabling option.

Note

To ensure if the passcode protection option is enabled/disabled, use \triangle and $\overline{}$ to toggle between the **ON** and **OFF** options. Display the desired option and press \bigcirc to confirm the selection.

The LCD returns to the EXIT screen.

5. Press ▲ or ▼ to select another user option, or press ○ to exit the user options and return to normal operation.

Note

If a passcode value is not selected or confirmed by pressing within 10 seconds, NO displays and the LCD returns to the EXIT screen.



Deactivation Passcode Protection

As a backup safety precaution, the deactivation passcode protection option can be enabled prevent unauthorized deactivation. A separate security passcode is required for this option and will be available to limited personnel only.

Note


The detector can be shipped with this option enabled permanently. This option can only be enabled at the factory and cannot be disabled by a customer.

The passcode must be entered every time the detector is deactivated.

User Manual

To deactivate the detector, complete the following:

If the detector is passcode protected to prevent deactivation, **OFF** displays briefly and then **PASS** immediately displays.



Press or to scroll to the required security passcode.
 Press to confirm the selection.

 \smile

The detector then deactivates.

Stealth Mode Option

The stealth (STLH) mode option ensures that the detector is undetected in situations that require concealment. This option disables the

- · audible alarms.
- · visual alarms, and
- backlight.

Only the vibrator alarm remains enabled.

Note


The detector is shipped with stealth mode disabled.

To enable/disable the stealth mode, complete the following:

From the EXIT screen, press ▲ or ▼ to scroll to the STLH option.



Press to select the option. The LCD flashes either ON or OFF.



3. Press ▲ or ▼ to toggle between the **ON/OFF** options. Ensure the desired option is displayed and press ○ to confirm the selection.

The LCD returns to the **EXIT** screen.

If stealth mode has been enabled, the screen displays **STLH** continually unless

- · functions are being performed,
- readings are not 0 ppm for toxics, or
- reading is not 20.9% vol for oxygen.

Note

The vibrator alarm is disabled at -20°C.

 Press ▲ or ▼ to scroll to a new user option or press ○ to exit and return to normal operation.

User Manual

Note

If the option is not confirmed by pressing within 10 seconds, the detector returns to normal operation.

Automatic Backlight Option

The automatic backlight (**BKLT**) option enables or disables the backlight of the detector. When enabled, the backlight automatically activates for 3 seconds whenever there is insufficient light to view the LCD.

Press any button to activate the backlight for 6 seconds.

Note

The detector is shipped with the automatic backlight option enabled. The backlight option is not available in the user options menu when stealth mode is enabled.

To enable/disable the automatic backlight, complete the following:

From the EXIT screen, press ▲ or ▼ to scroll to the BKLT option.



- Press
 to select the option. The LCD flashes either ON or OFF.
- 3. Press ▲ or ▼ to toggle between the **ON/OFF** options. Ensure the desired option is displayed and press to confirm the selection.

The LCD returns to the EXIT screen.

 Press ♠ or ▼ to scroll to a new user option or press ○ to exi and return to normal operation.

Note

If the option is not confirmed by pressing \(\) within 10 seconds, the detector returns to normal operation.

Latching Alarm Option

The latch alarm (LTCH) option ensures that an alarm persists until it is acknowledged by the user.

If enabled, during an alarm condition the latched alarms (LTCH) option causes the low and high gas alarms (audible, visual, and vibrator) to persist until the gas concentration is below the alarm setpoint and the alarms have been acknowledged by pressing ().

The audible alarm can be temporarily deactivated (press ()) for 30 seconds, but the LCD continues to display the peak concentration until the alarm condition no longer exists.

In stealth mode, the detector continues to vibrate until the alarm is acknowledged.

Note

The detector is shipped with the latching alarm option disabled.

User Manual

To enable/disable the latching alarm option, complete the following:

From the EXIT screen, press ▲ or ▼ to scroll to the LTCH option.



- Press to select the option. The LCD flashes either ON or OFF.
- 3. Press ▲ or ▼ to toggle between the **ON/OFF** options. Ensure the desired option is displayed and press to confirm the selection.

The LCD returns to the **EXIT** screen.

4. Press ▲ or ▼ to scroll to a new user option or press ○ to exit and return to normal operation.

Note

If the option is not confirmed by pressing \(\) within 10 seconds, the detector returns to normal operation.

Automatic Oxygen (O2) Calibration Option

Note

For oxygen detectors only.

If the automatic oxygen (O_2) calibration option is enabled, ensure the detector is activated in safe area in normal (20.9%) oxygen atmosphere.

This option enables/disables the automatic oxygen (O_2) calibration. The O_2 calibration begins automatically during start-up after the calibration due screen displays.

Note

The detector is shipped with the automatic O_2 calibration option disabled.

To enable/disable the automatic O₂ calibration option, complete the following:

From the EXIT screen of the user options menu, press ▲ or ▼
to scroll to the ACAL option.



- 2. Press to select this option. The LCD flashes either **ON** or **OFF**.
- 3. Press ▲ or ▼ to toggle between the **ON/OFF** options. Ensure the desired option is displayed and press to confirm the selection.

The LCD returns to the **EXIT** screen.

 Press ▲ or ▼ to scroll to a new user option or press ○ to exit and return to normal operation.

Note

If the option is not confirmed by pressing within 10 seconds, the detector returns to normal operation.

User Manual

Calibration Past Due Option

The calibration past due (**PAST**) option enables an automatic shutdown during start-up if the detector is past due for calibration.

Note

The detector is shipped with the calibration past due shutdown option disabled.

To calibrate a past due calibration detector, refer to <u>Calibration Past Due Enabled</u> in <u>Self-Test</u>.

To enable/disable the calibration past due automatic shutdown option, complete the following:

From the EXIT screen of the user options menu, press ▲ or ▼
to scroll to the PAST option.



- Press
 to select the option. The LCD flashes either ON or OFF.
- 3. Press ▲ or ▼ to toggle between the **ON/OFF** options. Ensure the desired option is displayed and press to confirm the selection.

The LCD returns to the **EXIT** screen.

4. Press ▲ or ▼ to scroll to a new user option or press ○ to exit and return to normal operation.

Note

If the option is not confirmed by pressing within 10 seconds, the detector returns to normal operation.

Languages

The LCD can display text in five different languages. Refer to the following language options.

Portuguese Option

The Portuguese (**PORT**) option enables the LCD to display text in Portuguese.

Note

If the multi-language option is included, the detector is shipped with English selected as the default language.

From the EXIT screen of the user options menu, press ▲ or ▼
to scroll to the PORT option.



2. Press O to select the option. The LCD then displays the Portuguese exit screen.



3. Press ▲ or ▼ to scroll to another user option or press ○ to exit and return to normal operation.

User Manual

Spanish Option

The Spanish (ESPA) option enables the LCD to display text in Spanish.

Note

If the multi-language option is included, the detector is shipped with English selected as the default language.

From the EXIT screen of the user options menu, press or to scroll to the ESPA option.



Press to select the option. The LCD then displays the Spanish exit screen.



3. Press ▲ or ▼ to scroll to another user option or press ○ to exit and return to normal operation.

German Option

The German (**DEUT**) option enables the LCD to display text in German.

Note

If the multi-language option is included, the detector is shipped with English selected as the default language.

From the EXIT screen of the user options menu, press ▲ or ▼
to scroll to the DEUT option.



Press to select the option. The LCD then displays the German exit screen.



3. Press ▲ or ▼ to scroll to a new user option or press ○ to exit and return to normal operation.

French Option

The French (**FRAN**) option enables the LCD to display text in French.

Note

If the multi-language option is included, the detector is shipped with English selected as the default language.

From the EXIT screen of the user options menu, press ▲ or ▼
to scroll to the FRAN option.



Press to select the option. The LCD then displays the French exit screen.



3. Press ▲ or ▼ to scroll to another user option or press ○ to exit and return to normal operation.

English Option

The English (**ENGL**) option enables the LCD screens to display text in English.

Note

If the multi-language option is included, the detector is shipped with English selected as the default language.

1. From the **EXIT** screen of the user options menu, press ▲ or ▼ to scroll to the **ENGL** option.



Press to select the option. The LCD then displays the English exit screen.



3. Press ▲ or ▼ to scroll to another user option or press ○ to exit and return to normal operation.

Datalogger Sampling Rate Option

The datalogger sampling rate (**RATE**) option defines how often the detector records a datalog. The datalogger sampling rate ranges from 1 to 60 seconds.

Note

The detector is shipped with a datalogging sampling rate of **5** seconds.

To adjust the datalogger sampling rate, complete the following:

From the EXIT screen of the user options menu, press ▲ and
 to scroll to the RATE option.



2. Press () to select the option and display the sample rate screen.



User Manual

4. Press ▲ or ▼ to scroll to another user option or press ○ to exit and return to normal operation.

Note

If a datalogging sample rate value is not selected or confirmed by pressing \(\circ\) within 10 seconds, **NO** displays and the LCD displays the **EXIT** screen.



Data Transfer Option

The data transfer (**SEND**) option transfers the datalog/event log information from the detector to a PC.

Note


An IR DataLink (or other BW accessory) is required to transfer the data from the detector to a PC.

To transfer data, complete the following:

- Connect the IR DataLink (or other BW accessory) to the detector and the PC
 - Refer to the IR DataLink Instruction Sheet.
- 2. From the **EXIT** screen of the user options menu, press ▲ or ▼ to scroll to the **SEND** option.



- 3. Press to select the option and to access the data transfer option screens.
- 4. Select one of the following options to transfer data.



When the data transfer is complete, the detector beeps and vibrates, and the LEDs flash. The LCD displays the EXIT screen.

User Manual

LAST and ALL Transfers

If the **LAST** or **ALL** option is selected, the LCD displays a countdown and the data transmission icon to indicate that the detector is transferring data.



The number at the beginning of the countdown depends upon the amount of data to transfer.

EVNT Transfer

If the **EVNT** option is selected, the event logs transfer immediately and the LCD displays the **EXIT** screen.

Unsuccessful Transfer

If the connection between the detector and the IR DataLink is disconnected during a transfer, **FAIL** displays.



The LCD then displays the **EXIT** screen.

- 1. From the PC, save the previously transferred data to ensure that it will not be deleted.
- 2. Repeat steps #3-5 of the Data Transfer Option.
- From the detector, select LAST to automatically resume the transfer from where it stopped sending.

Or

Select ALL to transfer all of the data again.

Alarms

Table 6 describes detector alarms and corresponding screen. During an alarm condition, the detector activates the backlight and the LCD displays the current ambient gas reading.

To change the factory-set alarm setpoints, refer to <u>Calibration and Setting Alarm Setpoints</u>.

Table 6. Alarms

Alarm	Display	Alarm	Display
Slow beepSlow flashALARM flashesSlow vibrations	LOW ALARM	TWA Alarm: Slow beep Slow flash ALARM flashes Slow vibrations	TWA ALARIM
 High Alarm: Fast beep Fast flash ALARM flashes Fast vibrations 	HIGHALARM	STEL Alarm:	STEL ALARM
Sensor Alarm:	ALARM	Cow Battery Alarm: One beep and one flash every 5 seconds, and one quick vibration every minute (when confidence beep is disabled). If confidence beep is enabled no beeps, flashes, or vibrations LOW displays	LOW

Table 6. Alarms

Alarm	Display	Alarm	Display
Automatic Shutdown Alarm: (Low battery) • Eight beeps, flashes, and vibrations • LOW displays	LOW	Automatic Shutdown Alarm: (Calibration past) • Eight beeps, flashes, and vibrations	OFF
Automatic Shutdown Alarm: (After Automatic Shutdown Alarm) No beep No flash or vibrations displays for a short time	û	Confidence Beep: One beep every 5 seconds One quick vibration per minute Note If the detector enters low battery alarm, the confidence beep deactivates.	I-IPPM

Note

The high alarm and STEL alarm have the same priority. A high alarm and/or STEL alarm overrides a low alarm and/or TWA alarm. To check STEL and TWA alarms specifically, press and hold and simultaneously.

The vibrator alarm is disabled at -20°C.

The high and low alarms deactivate when the gas concentration is lower than the low alarm setpoint. If the alarms are set to latch, alarms persist until the gas concentration is below the alarm setpoint and the alarms have been acknowledged by pressing . The TWA and STEL alarms deactivate by clearing the TWA and STEL peak exposure. Refer to Clearing Gas Exposures.

Computed Gas Exposures

To avoid possible personal injury, do not deactivate the detector during a work shift. The detector automatically resets the TWA, STEL, and MAX gas exposures during start-up. If the detector is reactivated during a work shift, the new values will not reflect the entire work shift.

Table 7. Computed Gas Exposures

Gas Exposure	Description	
TWA	Time-weighted average based on an 8-hour workday.	
	Accumulated value.	
STEL	Short-term exposure limit (STEL) to gas based on a 15-minute period.	
	Accumulated value.	
MAX*	Maximum (MAX) concentration encountered during a work shift.	

^{*}For oxygen, it is the highest or the lowest value from 20.9% encountered.

Viewing Gas Exposures

Toxic Gases

Press and simultaneously. The LCD displays the TWA gas exposure first.



2. Then the LCD displays the STEL gas exposure.



3. Then the LCD displays the MAX gas exposure.



Oxygen

For oxygen detectors, press \bigcirc and \blacktriangle simultaneously to view both the maximum low and maximum high levels of oxygen exposure.



31

Gas Alarm Setpoints

Table 8 describes the gas setpoints that trigger the gas alarms.

Table 8. Gas Alarm Setpoints

Alarm	Condition	
Low alarm	Toxic gases: Ambient gas level above low alarm setpoint.	
	O ₂ : ambient gas level may be set to above or below 20.9%.	
High alarm	Toxic gases: ambient gas level above high alarm setpoint.	
	O ₂ : ambient gas level may be set to above or below 20.9%.	
TWA alarm	TWA above TWA alarm setpoint. (O ₂ : not applicable)	
STEL alarm	STEL above STEL alarm setpoint. (O ₂ : not applicable)	

Stopping a Gas Alarm

The low and high alarms deactivate when the ambient gas level returns to below the low alarm setpoint.

Note

If alarms are set to latch, the alarms deactivate after the gas concentration is lower than the low alarm setpoint and the alarms have been acknowledged by pressing ().

The TWA and STEL alarms can be stopped either by

 Clearing the MAX, TWA, and STEL peak exposures. Refer to Clearing Gas Exposures.

or

· deactivating the detector and reactivating it again.

If the detector is passcode protected to prevent deactivation, refer to Deactivation Passcode Protection.

∧ Caution

Follow all safety procedures as defined by your employer. Confirm with your supervisor before clearing TWA and STEL alarms.

User Manual

Clearing Gas Exposures

The peak gas exposures automatically clear after deactivating the detector.

To clear the MAX, TWA, and STEL peak exposure readings immediately, press and hold \bigcirc for 6 seconds. The detector beeps and vibrates two times to confirm that the exposures have been cleared.

⚠ Caution

Follow all safety procedures as defined by your employer. Confirm with your supervisor before clearing TWA and STEL alarms.

Resetting Gas Alarm Setpoints

Note

Standard factory alarm setpoints vary by region.

Table 9 lists the factory alarm setpoints.

To change the factory alarm setpoints, refer to <u>Calibration and Setting Alarm Setpoints</u>.

Note

To disable an alarm, set the alarm setpoint to 0.

The ETO sensor is extremely cross sensitive and it responds strongly to CO.

Table 9. Factory Alarm Setpoints

Gas	TWA	STEL	Low	High
O ₂	N/A	N/A	19.5% vol.	22.5% vol.
CO (low H ₂)	35 ppm	200 ppm	35 ppm	200 ppm
СО	35 ppm	200 ppm	35 ppm	200 ppm
H ₂ S (high range)	10 ppm	15 ppm	10 ppm	15 ppm
H ₂ S	10 ppm	15 ppm	10 ppm	15 ppm
PH ₃	0.3 ppm	1.0 ppm	0.3 ppm	1.0 ppm
SO ₂	2.0 ppm	5.0 ppm	2.0 ppm	5.0 ppm
Cl ₂	0.5 ppm	1.0 ppm	0.5 ppm	1.0 ppm
NH₃	25 ppm	35 ppm	25 ppm	50 ppm
NH ₃ (high range)	25 ppm	35 ppm	25 ppm	50 ppm
NO ₂	2.0 ppm	5.0 ppm	2.0 ppm	5.0 ppm
HCN	4.7 ppm	10.0 ppm	4.7 ppm	10.0 ppm
ETO	1.0 ppm	5.0 ppm	1.0 ppm	5.0 ppm
CIO ₂	0.1 ppm	0.3 ppm	0.1 ppm	0.3 ppm
O ₃	0.10 ppm	0.10 ppm	0.10 ppm	0.20 ppm
NO	25 ppm	25 ppm	25 ppm	25 ppm

Sensor Alarm

The detector tests for a missing or defective sensor during the activation self-test. Refer to <u>Troubleshooting</u>.

User Manual

Low Battery Alarm

If a low battery alarm occurs, follow your company's safety procedures.

The detector tests the battery upon activation and continuously thereafter. If the battery voltage is low, the detector activates the low battery alarm.

The low battery alarm continues until the battery is replaced or the battery power is almost depleted. If the battery voltage drops too low, the detector automatically deactivates.

Note

If the detector enters low battery alarm, the confidence beep deactivates.

Replace the battery in only a safe area, free of hazardous gas.

Automatic Shutdown Alarm

There are two situations when an automatic shutdown alarm occurs.

 If the battery voltage is in immediate danger of falling below the minimum operating voltage, the detector beeps and vibrates eight times, and the LEDs flashes eight times. After 3 seconds, the LCD deactivates and the detector deactivates.

The LCD periodically displays $\hat{\mathbf{Q}}$ until the battery power is depleted.

To replace the battery, refer to Replacing the Battery or Sensor.

Note

The low battery alarm continues for approximately 30 minutes before an automatic shutdown occurs.

If the calibration past due user option is enabled and the detector is past the calibration due date, the detector automatically deactivates.

Calibration and Setting Alarm Setpoints

Guidelines

When calibrating the detector, adhere to the following guidelines.

Recommended gas mixture:

O₂: clean air, 20.9% vol.

CO (low H₂ sensitivity): 50 to 500 ppm balance N₂

CO: 50 to 500 ppm balance N₂

H₂S (high range): 10 to 100 ppm balance N₂

H₂S: 10 to 100 ppm balance N₂

PH₃: 1 to 5 ppm balance N₂

SO₂: 10 to 50 ppm balance N₂

Cl₂: 3 to 25 ppm balance N₂

NH₃: 20 to 100 ppm balance N₂

NH₃: (high range) 20 to 100 ppm balance N₂

NO₂: 5 to 50 ppm balance N₂

HCN: 5 to 20 ppm balance N₂

ETO: 5 to 50 ppm balance N₂

CIO₂: 0.1 to 1.0 ppm balance N₂

 O_3 : 0.1 to 1.0 ppm balance N_2

NO: 10 to 250 ppm balance N₂

- Before operating an ETO detector, allow the detector to stabilize at the temperature it will be operating in. After the detector has stabilized, zero the detector.
- It is necessary to periodically re-zero the ETO detector.
- To ensure accurate calibration, BW recommends using a premium-grade calibration gas approved by the National Institute of Standards and Technology (NIST).
- Do not use a gas cylinder beyond its expiration date.

- Before calibrating a new NO or ETO sensor, allow the sensor to stabilize for 2 hours in a safe area that is free of hazardous gas.
- Calibrate a new sensor before use. Allow the sensor to stabilize before starting calibration (used: 60 seconds; new: 5 minutes).
- Calibrate the detector at least once every 180 days (for HCN detectors calibrate at least once every 90 days), depending upon use and sensor exposure to poisons and contaminants.
- Calibrate the detector if the ambient gas display varies at start-up.
- It is best to calibrate the sensor before changing the alarm setpoints.
- Calibrate only in a safe area that is free of hazardous gas.
- To disable an alarm, set the alarm setpoint to zero.
- If a certified calibration is required, contact BW Technologies by Honeywell.

Note

A generator must be used to calibrate O_3 and CIO_2 GasAlert Extreme sensors.

Test Cap

The calibration cap and hose are shipped with the detector for calibration.

Refer to Table 10 and Figure 3 for installation information.

Note

Only use the calibration cap during calibration.

Table 10. Test Cap

ltem	Description
1	Test cap
2	Hose
3	Regulator
4	Gas cylinder




Figure 3. Test Cap

User Manual

Calibration

Calibration requires several steps, some of which can be bypassed. A note is added to each option that can be bypassed.

Start Calibration

Note

To quit calibration at any point, press ①. The detector retains any saved values and the detector beeps and vibrates four times before returning to normal operation.

Calibrate O₂ in clean air only.

 To enter calibration, in a safe area free of hazardous gas, press and hold and simultaneously as the detector beeps and vibrates four times, and the LEDs flash four times.



After the **CAL**. screen displays, the detector beeps one time and the auto zero screen displays.

Auto Zero

The auto zero function automatically zeroes the detector.



 The LCD flashes while the detector automatically zeroes the sensor. When the auto zero is complete the detector beeps twice.

Note

Do not apply the calibration gas until the LCD displays the flashing gas cylinder icon; otherwise, the detector auto zero will fail.

Auto Zero Fail

If the sensor fails auto zero, the following screen displays.



The detector then bypasses the sensor span and automatically proceeds to the alarm setpoints.

- Press to exit the alarm setpoint screens and to return to normal operation.
- Restart the calibration procedures in a safe area that is free of hazardous gas. If auto zero fails a second time, deactivate and then reactivate the detector to test the sensors.
- If the auto zero is successful and the passcode protection is disabled, the detector automatically proceeds to the auto span function.

User Manual

Passcode Protected

After a successful auto zero, and if the passcode protected option is enabled, **PASS** displays. The passcode is required to access the auto span and alarm setpoint functions.



Press ▲ or ▼ to scroll to the required passcode and press ○ to confirm. For additional information, refer to Passcode Protection Option.

If the correct code is confirmed by pressing \bigcirc within 10 seconds, the detector beeps twice and automatically proceeds to the set span screen.



If the passcode is not confirmed within 10 seconds or the passcode is incorrect, **NO** displays.



The detector then beeps four times and automatically returns to normal operation.

Set Span

Note

To bypass the set span function, press () to automatically proceed to the span screen.

The set span function inputs a new calibration gas concentration value.



4. Set SPAN flashes.

Press or to scroll to the required gas concentration. The detector value must match the concentration value on the gas cylinder.

Note

If a new value is selected but not confirmed within 10 seconds by pressing \bigcirc , the detector rejects the new value and **NO** displays. The detector beeps six times and retains the original value. The detector automatically proceeds to the span screen.



5. Press to save the new value and proceed to the span screen.



Span

Note

To bypass the span function, press \bigcirc to automatically proceed to the alarm setpoint screens. If the span is bypassed, the calibration due date cannot be changed.

Verify that the calibration gas being used matches the span concentration values that are defined for the detector. For more information, refer to <u>Calibration Guidelines</u>.

6. The set span screen displays a flashing •••.

Note

The flashing \blacksquare_{\bullet} does not display for oxygen (O_2) detectors.

- 7. Apply the calibration gas.
- 8. Apply gas to the sensor at a flow rate of 500 ml/min. (for NH₃, Cl₂, and ETO: 1000 ml/min.)

The gas readings change as gas is applied to the sensor. When the detector senses a sufficient concentration of gas (approximately 30 seconds), the detector beeps once.

- 9. The detector then begins spanning the sensor as follows:
 - NH₃, Cl₂, ClO₂, O₃, and ETO: 5 minutes to span
 - O₂: 30 seconds to span
 - other gases: 2 minutes (approximately) to span.

The detector beeps three times when the span is complete.

Successful Span

If the span is successful, the LCD automatically displays the calibration due date screen.

Unsuccessful Span

If the detector fails to span a sensor successfully, FAIL displays.



The detector vibrates and beeps, and the LEDs flash. Then the detector automatically proceeds to the alarm setpoint screens.

If the span fails confirm that

- · gas is being applied to the sensor,
- the sensor is detecting a sufficient gas concentration within 30 seconds, and
- the gas concentration has not dropped significantly during the 2-minute span.

If the span is still unsuccessful, use a new gas cylinder.

If the span continues to be unsuccessful, replace the sensor. Refer to Replacing the Battery or Sensor.

User Manual

Setting the Calibration Due Date

After a successful calibration, the LCD displays the **CAL. DUE** screens and the number of days remaining before the next calibration.







Note

To bypass the calibration due notification, press . The detector automatically proceeds to the TWA alarm setpoint.

BW Technologies by Honeywell recommends that the detector be calibrated every 180 days (6 months). The detector is shipped with the factory default setting of 180 days.

- 10. Press ▲ or ▼ to scroll to the required value (1 to 365).
- 11. Press to save the new value and automatically proceed to the TWA alarm setpoint screen.

Note

If a new value is selected but not confirmed within 10 seconds by pressing \bigcirc , the detector automatically retains the original value and **NO** displays. The detector proceeds to the TWA alarm setpoint.



Setting the TWA Alarm Setpoint

Note

To bypass and retain the current TWA alarm setpoint value, press . The detector automatically proceeds to the STEL alarm setpoint.

When the **CAL**. **DUE** function has been completed, the **Set TWA** alarm setpoint screen automatically displays.



- 12. Press ▲ or ▼ to scroll to the required value.
- 13. Press to save the new value and proceed to the STEL alarm setpoint.

Note

If a new value is selected but not confirmed within 10 seconds by pressing \bigcirc , the detector automatically retains the original value and **NO** displays. The detector proceeds to the STEL alarm setpoint.



User Manual

Setting the STEL Alarm Setpoint

Note

To bypass and retain the current STEL alarm setpoint value, press . The detector automatically proceeds to the low alarm setpoint.

When the TWA alarm setpoint value has been changed or bypassed, the **Set STEL** alarm setpoint screen displays.



- 14. Press ▲ or ▼ to scroll to the required value.
- 15. Press to save the new value and proceed to the low alarm setpoint.

Note

If a new value is selected but not confirmed within 10 seconds by pressing \bigcirc , the detector automatically retains the original value and **NO** displays. The detector proceeds to the low alarm setpoint.



Setting the Low Alarm Setpoint

Note

To bypass and retain the current low alarm setpoint value, press . The detector automatically proceeds to the high alarm setpoint.

When the STEL alarm setpoint value has been changed or bypassed, the **Set LOW** alarm setpoint screen displays.



- 16. Press ▲ or ▼ to scroll to the required value.
- 17. Press () to save the new value and proceed to the high alarm setpoint.

Note

If a new value is selected but not confirmed within 10 seconds by pressing \bigcirc , the detector automatically retains the original value and **NO** displays. The detector proceeds to the high alarm setpoint.



User Manual

Setting the High Alarm Setpoint

Note

To bypass and retain the current high alarm setpoint value, press (). The detector then returns to the normal operation.

When the low alarm setpoint value has been changed or bypassed, the **Set HIGH** alarm setpoint screen displays.



- 18. Press ▲ or ▼ to scroll to the required value.
- 19. Press () to save the new value and return to normal operation.

Note

If a new value is selected but not confirmed within 10 seconds by pressing (), the detector automatically retains the original value and **NO** displays. The detector proceeds to normal operation.



When calibration is complete, the detector beeps and vibrates four times, and the LEDs flash four times before returning to normal operation.

Verification

- After calibration is complete and the detector is in normal operation, verify calibration by using a gas cylinder other than the one used for calibration.
- The gas concentration should not exceed the sensor's detection range. Confirm that the LCD displays the expected concentration values.
- To ensure that the reading is accurate, apply the verification gas for the same period of time as was applied to the sensor when it was calibrated.

Example: SO₂ span time was 2 minutes therefore, apply verification gas for 2 minutes.

Datalog and Event Log

The GasAlert Extreme datalogger version allows the detector to record various information so a report can be compiled.

Datalog

Datalog sampling rate is defined in the detector user options. To set the sample rate, refer to <u>Datalogger Sampling Rate Option</u>.

The following information is recorded in a datalog:

- · Date and time
- · Detector serial number
- Type of gas the detector monitors
- · Current gas reading
- · Sensor status
- Detector status
- · Passcode protect enabled/disabled
- STEL period setting (fixed to a 15 minute period)
- Confidence beep enabled/disabled
- Automatic backlight enabled/disabled
- Stealth mode enabled/disabled
- · Latching alarm enabled/disabled
- Calibration past due user option enabled/disabled
- Language the detector is set to display

Event Log

Event log information is recorded when an event (i.e., an alarm) occurs. The following information is recorded in an event log:

- Detector serial number
- Type of exposure the detector that occurred
- · Time the alarm started and ended
- Peak exposure of the alarm

Maintenance

To maintain the detector in good operating condition, perform the following basic maintenance as required:

- Calibrate, bump test, and inspect the detector at regular intervals.
- Maintain an operations log of all maintenance, calibrations, bump tests, and alarm events.
- Clean the exterior with a soft damp cloth. Do not use solvents, soaps, or polishes.
- Do not immerse the detector in liquids.

Table 11. Replacing the Battery or Sensor

Item	Description
1	Rear shell machine screws (4)
2	Rear shell
3	Battery
4	PCB machine screws (2)
5	PCB
6	Sensor
7	Sensor screen
8	Front shell




Figure 4. Replacing the Battery or Sensor




Figure 5. Rear Shell Seal Table 12. Rear Shell Seal

Item	Description
1	Seal




Figure 6. Front Shell Seal Table 13. Front Shell Seal

Item	Description
1	Seal

Replacing the Battery or Sensor

Marning

To avoid possible personal injury, adhere to the following:

- Replace the battery in a safe area, free of hazardous gas immediately when the detector enters low battery alarm.
- Use only the Panasonic CR-2PE/BN battery.
- Use only the sensor specifically designed for the GasAlert Extreme model. Otherwise, the detector will not monitor the target gas. Refer to Replacement Parts and Accessories.
- After replacing a sensor, allow the new sensor 5 minutes to stabilize before use. For an ETO or NO sensor, allow the new sensor 2 hours to stabilize before use.
- Do not expose a sensor to vapors of organic solvents such as paint fumes or organic solvents.

Note

When the battery is removed from the detector, the clock reverts back to the default value. Refer to <u>Clock Option</u>.

To preserve the life of the battery, deactivate the detector when not in use.

For additional information regarding problems caused by a sensor requiring calibration or replacement, refer to <u>Troubleshooting</u>.

Replacing the Battery

To replace a battery, complete the following. Refer to Figures 4, 5, and 6 and Tables 11, 12, and 13. Replace the battery in a safe area, free of hazardous gas.

- 1. Deactivate the detector.
- Remove the four machine screws on the rear shell and remove the rear shell.
- 3. Remove the battery.

Warning

This instrument contains a lithium battery. Do not mix with the solid waste stream. Spent batteries should be disposed of by a qualified recycler or hazardous materials handler.

- 4. Insert the new battery.
- Re-assemble the detector. When assembling the detector be aware of the following:
 - Clean the seal on the front and rear shells with a soft damp clean cloth. Do not use solvents, soaps, or polishes. Refer to Figures 5 and 6.
 - Ensure the front and rear shells are properly aligned to ensure a proper environmental seal.
 - Torque the screws to 3-4 in-lbs in a crisscross pattern to ensure a proper environmental seal. Do not overtighten.

Replacing the Sensor

To replace a sensor, complete the following. Refer to Figures 4, 5, and 6 and Tables 11, 12, and 13.

- 1. Deactivate the detector.
- Remove the four machine screws on the rear shell and remove the rear shell.
- Remove the two machine screws from the PCB.

- 4. Remove the PCB.
- Replace the sensor.

Gently rock the sensor back and forth to remove a tightly held sensor.

Note

Allow the new sensor 5 minutes to stabilize before use. For a new ETO or NO sensor, allow the new sensor 2 hours to stabilize before use.

- Re-assemble the detector. When assembling the detector be aware of the following:
 - Clean the seal on the front and rear shells with a soft damp clean cloth. Do not use solvents, soaps, or polishes. Refer to Figures 5 and 6.
 - Ensure the front and rear shells are properly aligned to ensure a proper environmental seal.
 - Torque the screws to 3-4 in-lbs in a crisscross pattern to ensure a proper environmental seal. Do not overtighten.

Cleaning a Sensor Screen

Clean or replace the sensor screen as required. If replacement sensor screens are required, refer to Replacement Parts and Accessories.

To clean a removed sensor screen, complete the following. Refer to Figures 4, 5, and 6 and Tables 11, 12, and 13.

- 1. Deactivate the detector.
- Remove the four machine screws on the rear shell and remove the rear shell.

- 3. Remove the two machine screws from the PCB.
- 4. Remove the PCB. Place the PCB on a clean surface.
- Remove the screen.
- Using a soft, clean brush, wash the screen with clean, warm water.
- Insert the sensor screen with the shiny side facing the sensor grill.

Note

Ensure the screen is dry before inserting back into the detector.

- 8. Re-assemble the detector. When assembling the detector be aware of the following:
 - When inserting the sensor screen back into the detector, ensure the sensor screen is inserted with the shiny side facing the sensor grill.
 - Clean the seal around the edge of the front and rear shells with a soft damp clean cloth. Do not use solvents, soaps, or polishes. Refer to Figures 5 and 6.
 - Ensure the front and rear shells are properly aligned to guarantee a proper environmental seal.
 - Torque the screws to 3-4 in-lbs in a crisscross pattern to ensure a proper environmental seal. Do not overtighten.

GasAlert Extreme User Manual

Clearing a Sensor

Each sensor has a high degree of resistance to common vapors and gases. To clear a sensor, move the detector to a clean environment and wait 10 to 30 minutes.

Note

Do not expose a sensor to the vapors of inorganic solvents, such as paint fumes or organic solvents.

WEEE Directive and Battery Directive

Failure to comply with the following battery removal and disposal instructions may result in battery shorting, battery leakage, and/or other damage. Ensure a qualified technician completes the following procedures.

Removal and Disposal of the Battery

Only a qualified technician should complete the following procedure. Refer to Figure 4. for parts of the detector.

Dispose of the battery according to local laws.

- 1. Press and hold () to deactivate the detector.
- Remove the four machine screws on the rear shell and remove the rear shell.
- 3. Remove the two machine screws from the PCB.
- Remove the PCB. Place the PCB on a clean, non-conductive surface.
- 5. Remove the battery.
- 6. Dispose of the battery according to local laws.

Troubleshooting

If a problem occurs, refer to the solutions provided in Table 14. If the problem persists, contact <u>BW Technologies by Honeywell</u>.

Table 14. Troubleshooting Tips

Problem	Possible Cause	Solution
The detector does not activate.	No battery	Install a battery. Refer to Replacing the Battery or Sensor.
	Depleted battery	Replace the battery. Refer to Replacing the Battery or Sensor.
	Damaged or defective detector	Contact <u>BW Technologies by Honeywell</u> .
	Reversed battery	Reinstall the battery correctly.
The detector enters alarm mode immediately when it is activated.	Sensor needs to stabilize	Used sensor: wait 60 seconds. New sensor: wait 5 minutes.
	Low battery alarm	Replace the battery. Refer to Replacing the Battery or Sensor.
	Sensor alarm	Replace the sensor. Refer to Replacing the Battery or Sensor.
The start up self-test fails during one of the checks.	General fault	Contact BW Technologies by Honeywell.
	Alarm setpoints are incorrect	Reset the alarm setpoints. Refer to Resetting Gas Alarm Setpoints.

Table 14. Troubleshooting Tips

Problem	Possible Cause	Solution
The detector does not display a normal ambient gas reading after the activation	Target gas is present	Detector is operating properly. Use caution in suspect areas.
self-test.	Detector requires calibration	Calibrate the detector. Refer to <u>Calibration</u> and <u>Setting Alarm Setpoints</u> .
	Sensor not stabilized	Used sensor: wait 60 seconds. New sensor: wait 5 minutes.
The detector does not respond to the push-buttons.	Battery is depleted	Replace the battery. Refer to Replacing the Battery or Sensor.
	Detector is performing operations that does not require user input	Pushbutton operation restores automatically when the operation ends.
The detector does not accurately measure the gas.	Detector requires calibration	Calibrate the sensor. Refer to <u>Calibration</u> and <u>Setting Alarm Setpoints</u> .
	Detector is colder/hotter than ambient gas	Allow the detector to acquire ambient temperature before use.
	Sensor screen is blocked	Clean the sensor screen. Refer to <u>Cleaning</u> a <u>Sensor Screen</u> .
The detector does not enter alarm mode.	Alarm setpoint(s) are set incorrectly	Reset the alarm setpoints. Refer to Resetting Gas Alarm Setpoints.
	Alarm setpoint(s) set to zero	Reset the alarm setpoints. Refer to Resetting Gas Alarm Setpoints.
	Detector is in calibration mode	Complete calibration. Refer to <u>Calibration</u> and <u>Setting Alarm Setpoints</u> .

Table 14. Troubleshooting Tips

Problem	Possible Cause	Solution
The detector intermittently enters alarm mode without apparent reason.	Ambient gas levels are near alarm setpoint or the sensor is exposed to a puff of the target gas	Detector is operating normally. Use caution in suspect areas. Check MAX gas exposure reading.
	Alarms set incorrectly.	Reset the alarm setpoints. Refer to <u>Calibration and Setting Alarm Setpoints</u> .
	Missing or faulty sensor	Replace the sensor. Refer to Replacing the Battery or Sensor.
The detector automatically deactivates.	Automatic shutdown feature activated due to depleted battery	Replace the battery. Refer to < Replacing the Battery or Sensor.
Detector does not auto zero or calibrate.	Sensor may be expired	Change the sensor.
O ₂ sensor reading is erratic.	Sensor may be expired	Change the sensor.

Replacement Parts and Accessories A Warning

To avoid personal injury or damage to the detector, use only specified replacement parts.

Table 15. Replacement Parts and Accessories

Model No.	Description	Qty
SR-X10	Replacement O ₂ sensor	1
PS-RM04H	Replacement CO sensor (low H2 sensitivity)	1
PS-RM04	Replacement CO sensor	1
PS-RH04S	Replacement H ₂ S sensor	1
SR-P04	Replacement PH ₃ sensor	1
PS-RS04	Replacement SO ₂ sensor	1
PS-RC10	Replacement Cl ₂ sensor	1
SR-A04	Replacement NH ₃ sensor	1
SR-A204	Replacement NH ₃ sensor (high range)	1
PS-RD04	Replacement NO ₂ sensor	1
PS-RZ10	Replacement HCN sensor	1
SR-E04	Replacement ETO sensor	1
SR-V04	Replacement CIO ₂ sensor	1
SR-G04	Replacement O ₃ sensor	1
SR-N04	Replacement NO sensor	1
GAXT-SS	Sensor screens	10

Model No.	Description	Qty
GA-TC-1	Test cap and hose	1
GA-BALERT	Concussion-proof boot	1
GA-BSPLASH-K	Splash guard filters	5
GA-HC-1	Hard hat clip	1
GA-AG-1	Alligator clip (non-conductive)	1
GA-AG-2	Alligator clip (stainless-steel)	1
GA-CL-1	Belt clip with screw	1
GA-NS-1	Neckstrap with safety release	1
GA-LY-1	Short strap (6 in./15.2 cm)	1
GA-USB2	IR DataLink	1

Specifications

Instrument dimensions: 2.8 x 5.0 x 9.5 cm

(1.1 x 2.0 x 3.75 in.)

Weight: 82 g (2.9 oz.)

Operating temperature:

 H_2S , SO_2 , HCN: -40°C to +50°C (-40°F to +122°F)

CO: -30°C to +50°C (-22°F to +122°F)

 NH_{3} , NH_{3} (high range): -20°C to +40°C (-4°F to +104°F)

Other gases: -20°C to +50°C (-4°F to +122°F)

Operating humidity:

 $\mathsf{CO},\,\mathsf{H}_2\mathsf{S},\,\mathsf{SO}_2,\,\mathsf{CI}_2,\,\mathsf{HCN},\,\mathsf{NO}_2,\,\mathsf{NH}_3,\,\mathsf{PH}_3,\,\mathsf{ETO},\,\mathsf{NO},\,\mathsf{O}_3\!.$

15% to 90% relative humidity (non-condensing)

Cl2: 10% to 95% relative humidity (non-condensing)

CIO2: 15% to 95% relative humidity (non-condensing)

O2: 0% to 99% relative humidity (non-condensing)

Detection range:

GasAlert Extreme O₂: 0 – 30.0% vol. (0.1% vol. increments)

GasAlert Extreme CO: 0 – 1000 ppm (1 ppm increments)

GasAlert Extreme CO (low H₂ sensitivity): 0 – 1000 ppm

(1 ppm increments)

GasAlert Extreme H₂S: 0 – 100 ppm (1 ppm increments)

GasAlert Extreme H₂S (high range): 0 – 500 ppm (1 ppm increments)

GasAlert Extreme PH_3 : 0 – 5.0 ppm (0.1 ppm increments)

GasAlert Extreme SO₂: 0 – 100.0 ppm (0.1 ppm increments)

GasAlert Extreme Cl₂: 0 – 50.0 ppm (0.1 ppm increments)

GasAlert Extreme NH₃: 0 – 100 ppm (1 ppm increments)

GasAlert Extreme NH₃: (high range) 0 – 400 ppm (1 ppm increments)

GasAlert Extreme NO₂: 0 – 100.0 ppm (0.1 ppm increments)

GasAlert Extreme HCN: 0 – 30.0 ppm (0.1 ppm increments)

GasAlert Extreme ETO: 0 – 100.0 ppm (0.1 ppm increments)

GasAlert Extreme ClO₂: 0 – 1.00 ppm (0.01 ppm increments)

GasAlert Extreme O_3 : 0 – 1.00 ppm (0.01 ppm increments)

GasAlert Extreme NO: 0 – 250 ppm (1 ppm increments)

Sensor type: Plug-in electrochemical cells

Calibration: Auto zero, set span, and span sensor

Alarm conditions: TWA alarm, STEL alarm, low alarm, high alarm, sensor alarm, low battery alarm, confidence beep, automatic shutdown

alarm.

Audible alarm: 95 dB at 0.3 m (1 ft.) typical

Visual alarm: Red light-emitting diode (LED)

Display: Alpha-numeric liquid crystal display (LCD)

Backlight: Automatically activates for 3 seconds whenever there is insufficient light to view the display and during alarm conditions. Any

pushbutton reactivates the backlight for 6 seconds.

Self-test: Initiated upon activation

Battery test: Every 0.5 seconds

Battery: Panasonic CR-2PE/BN battery

Warranty: 2 years including sensors.

Approvals:

Classified by UL to both U.S. and Canadian Standards as Intrinsically

Safe for Class I, Division 1, Group A, B, C, D European Explosives Protection EEx ia IIC

OF 0500 © # 4 O DENIKO 04 ATEV 00 00000

CE 0539 (II 1 G DEMKO 04 ATEX 03 36363

IECEx

ABS Type Approved: VA-348-169-X

KTL: 12-KB4BO-0199X¹

EAC Certificate: RU C-GB.ΓБ05.B.01115

User Manual

Inmetro: EX ia IIC T4 Ga DNV 12.0137 X

Ratings and Certifications

The GasAlertExtreme is in conformity with the following standards:

UL 913 8th Edition

CSA C22.2 No. 157-92:2012

EN 60079-0:2012 +A11:2013

EN 60079-11:2012

EN 60079-26:2007

IEC 60079-0:2011

IEC 60079-11:2011

IEC 60079-26:2006

General Specifications for Datalogger Units

Storage: Maximum of 8 months of data at 5 second intervals (based on a normal workweek).

Memory Type: Wrap-around memory ensures most recent data is always saved.

Sample Rate: One reading every 5 seconds (standard)

Data Recorded: All sensor readings, all alarm conditions, calibrations, event flags, battery status, sensor status, confidence beep activation, and detector status along with the time and date and the detector serial number.

Indicators: Icon advising datalogger is operating normally Transfer Accessory: IR DataLink or other BW accessory

Support:

BW Excel Datalog Manager (EDM): This software organizes GasAlert Extreme datalog and event log files into a readable report

Fleet Manager CD Support: This software organizes GasAlert Extreme datalog and event log files into a readable report

http://www.honeywellanalytics.com/~/media/honeywell-analytics/ products/gasalert-extreme/documents/certifications/koreanexcertificate bw gaextreme 12kb4bo0199x-(2).pdf?la=en

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules and ICES-003 Canadian EMI requirements. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that radio interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

User Manual

